Aactive HA

0.2%

10ml

Sodium Hyaluronate 0.2% Eye Drops

Instructions for Use

Composition:

Sodium Hyaluronate .....0.2% w/v

Sodium Perborate (As Preservative).....0.028% w/v

Aqueous buffered vehicle.....Q.S.

Intended use:

Temporary dry eye relief, burning and ocular fatigue and other minor complaints of no pathological significance induced, for example, by dust, smoke, dry heat, air conditioning, wind, cold, extended computer screen use of contact lens wear (rigid or soft).

Mode of Action:

Sodium Hyaluronate binds to fibronectin and accelerates the adhesion and extension of epithelial cells. Sodium hyaluronate also has an excellent water-holding property because each sodium hyaluronate molecule can retain many H_2O molecules.

Instructions for use:

Before opening a new bottle of Sodium Hyaluronate 0.2% Eye Drops, please check that the safety ring on the cap of the bottle is unbroken.

- Wash and dry your hands well before use.
- Rotate the cap anticlockwise to break the seal.
- Remove the cap.
- Tilt your head back and look up.
- Gently pull down the lower lid of the eye to form a small pocket between your eyelid and your eye.
- Hold the bottle between your thumb and middle finger of your other hand and with your index finger apply moderate pressure to the base of the eye drop bottle to release one drop of sodium hyaluronate into the pocket.
- Blink your eye a few times.
- If you think that you have missed the eye, then insert another drop.
- Carefully wipe away any excess sodium hyaluronate eye drops around the eyelid with a lint free tissue.
- Repeat the process for the other eye, if required.
- Replace the cap on the container, trying not to touch the applicator tip with anything, including the eye or your fingers.
- The usual dosage is one or two drops to be applied 2 to 4 times a day as needed.

WARNING:

- Do not use it if the safety ring on the cap of the bottle is broken.
- Do not touch the tip of the bottle with your eye or any surface.
- Not for injection. For external use only.
- Do not use it if the solution changes colour or becomes cloudy.
- Discard if the seal is broken.
- Discard the bottle 28 days after first opening.
- Do not freeze.
- To avoid contamination, do not touch the tip of the container to any surface.
- Overuse may cause eye redness.
- If you are a contact lens wearer remove your lenses and do not replace them until 30 minutes after using the device.

POTENTIAL ADVERSE EVENTS:

- Irritation of the eye
- Transient mild burning sensation
- Blurring of vision
- Eye pain

Contraindications:

- Hypersensitivity to any of the ingredients, in the even of persisting eye irritation discontinues use and consults your doctor.
- Temporary blurred vision after installation. Do not drive or operate hazardous machinery unless vision is clear.

Storage Conditions:

- Keep out of reach of children.
- Store between 0^0C to 25^0 C.

Sr. No	Symbol	Meaning
01	REF	Symbol for “Reference Number”
02	LOT	Symbol for “Batch Code”
03		Symbol for “Date of Manufacture”
04		Symbol for “Use Before”
05		Symbol for “Consult instructions for use”
06	STERILE A	Symbol for “Sterile”
07		Symbol for “Manufacturer”
08		Symbol for “Do not use if package is damaged”
09		Symbol for “keep away from sunlight”
10		Symbol for “Keep Dry”
11	CE 2460	CE Mark with Notified Body Number
12	UK RP	United Kingdom Representative
13	EC REP	European Representative

Exclusively distributed by:
TriOn Pharma Ltd - UK
Cowplain PO8 8DL, UK
info@trionpharma.co.uk

EC REP
M/s. Cmc Medical Devices & Drugs S.L.
C/Horacio Lengo 18, CP29006, Malaga Spain.
E-mail: Info@cmcmedicaldevices.com
www :cmcmedicaldevices.com

UK RP
CMC Medical Devices LTD.
Office 32 19-21, Crawford Street,
London W1H 1PJ - UK